

CERTIFICATE OF ACCREDITATION

In terms of section 22(2) (b) of the Accreditation for Conformity Assessment, Calibration and Good Laboratory Practice Act, 2006 (Act 19 of 2006), read with sections 23(1), (2) and (3) of the said Act, I hereby certify that:-

CONCILIUM TECHNOLOGIES (PTY) LTD
Co. Reg. No.: 1999/013330/07
ELECTRICAL DC LOW FREQUENCY LABORATORY

Accreditation Number: **106**

is a South African National Accreditation System accredited Calibration Laboratory
provided that all SANAS conditions and requirements are complied with

This certificate is valid as per the scope as stated in the accompanying scope of accreditation
Annexure "A", bearing the above accreditation number for

DC LOW FREQUENCY METROLOGY

The facility is accredited in accordance with the recognised International Standard

ISO/IEC 17025:2017

The accreditation demonstrates technical competency for a defined scope and the operation of a
laboratory quality management system

While this certificate remains valid, the Accredited Facility named above is authorised to use the
relevant SANAS accreditation symbol to issue facility reports and/or certificates

Mr T Baleni
Acting Chief Executive Officer

Effective Date: 31 January 2022
Certificate Expires: 30 January 2027

ANNEXURE A

SCOPE OF ACCREDITATION

DC LOW FREQUENCY METROLOGY

Accreditation Number: 106

Permanent Address of Laboratory: Concilium Technologies (Pty) Ltd Electrical DC Low Frequency Laboratory 1 Standford Office Park 12 Bauhinia Street Highveld Technopark Centurion 0157		Technical Signatories: Mr BJH Bremmer Mr GD Schuster		
Postal Address: P O Box 67611 Highveld 0169		Nominated Representative: Mr BJH Bremmer		
Tel: (012) 678-9200 Fax: (012) 665-4160 E-mail: bart_bremmer@concilium.co.za		Issue No.: 15 Date of Issue: 31 January 2022 Expiry Date: 30 January 2027		
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT AND RANGE OF MEASURED QUANTITY	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD/ PROCEDURE
1	DC Voltage (up to 1100 V)			
1.1	DC voltage sources			
1.1.1	Single values			
	1,0 V	DC	1•10 ⁻⁶ •U	Direct comparison with a DC voltage standard.
	10 V	DC	1•10 ⁻⁶ •U	
1.1.2	Low values (<= 10 V)			
	0 V to 10 V	DC	1•10 ⁻⁶ •U + 0,3 µV	Direct measurement or comparison with a DC voltage standard.
1.1.3	Intermediate values (>10 V to 1 100 V)			
	10 V to 100 V	DC	3•10 ⁻⁶ •U	Direct measurement or comparison with a DC voltage Source or multifunction calibrator.
	100 V to 1 000 V	DC	5•10 ⁻⁶ •U	
1.2	DC voltage meters			
1.2.1	Very low values (<= 10 V)			
	0 mV to 10 V	DC	1•10 ⁻⁶ •U + 0,3 µV	Direct measurement or comparison with a DC Voltage source and short.
1.2.2	Intermediate values (> 1 mV to 1 100 V)			
	10 V to 100 V	DC	3•10 ⁻⁶ •U	Direct measurement or comparison with a DC voltage source or multifunction calibrator.
	100 V to 1 000 V	DC	5•10 ⁻⁶ •U	

Original Date of Accreditation: 1980

Page 1 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%

Accreditation Manager

ANNEXURE A

Accreditation No.: 106
Date of Issue: 31 January 2022
Expiry Date: 30 January 2027

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT AND RANGE OF MEASURED QUANTITY	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	METHOD / PROCEDURE
2	DC Resistance			
2.1	DC Resistance Standards and Sources			
2.1.1	Low values ($\leq 1\Omega$)			
	1 m Ω 10 m Ω 100 m Ω 1 Ω	DC DC DC DC	$2 \cdot 10^{-4} \cdot R$ $1 \cdot 10^{-4} \cdot R$ $1 \cdot 10^{-4} \cdot R$ $5 \cdot 10^{-5} \cdot R$	Direct substitution or current volt drop method.
2.1.2	Intermediate values ($>1\Omega$ to 1 MΩ)			
	10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω	DC DC DC DC DC DC	$5 \cdot 10^{-5} \cdot R$ $5 \cdot 10^{-5} \cdot R$ $5 \cdot 10^{-5} \cdot R$ $5 \cdot 10^{-5} \cdot R$ $5 \cdot 10^{-5} \cdot R$ $1 \cdot 10^{-4} \cdot R$	Direct substitution or current volt drop method.
2.1.3	High values ($>1\text{ M}\Omega$)			
	10 M Ω 100 M Ω 1 G Ω	DC DC DC	$2 \cdot 10^{-4} \cdot R$ $5 \cdot 10^{-4} \cdot R$ $1 \cdot 10^{-3} \cdot R$	Direct or substitution method.
2.2	DC Resistance Meters			
2.2.1	Low values ($\leq 1\Omega$)			
	1 m Ω 10 m Ω 100 m Ω 1 Ω	DC DC DC DC	$2 \cdot 10^{-4} \cdot R$ $1 \cdot 10^{-4} \cdot R$ $1 \cdot 10^{-4} \cdot R$ $5 \cdot 10^{-6} \cdot R$	Direct or substitution method.
2.2.2	Intermediate values ($>1\Omega$ to 1 GΩ)			
	10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω	DC DC DC DC DC DC DC DC	$5 \cdot 10^{-5} \cdot R$ $5 \cdot 10^{-5} \cdot R$ $5 \cdot 10^{-5} \cdot R$ $5 \cdot 10^{-6} \cdot R$ $5 \cdot 10^{-5} \cdot R$ $1 \cdot 10^{-4} \cdot R$ $2 \cdot 10^{-4} \cdot R$ $5 \cdot 10^{-5} \cdot R$	Direct or substitution method.
2.2.3	High values ($>1\text{ G}\Omega$)			
	1 G Ω 10 G Ω 100 G Ω	DC DC DC	$1 \cdot 10^{-3} \cdot R$ $2 \cdot 10^{-3} \cdot R$ $5 \cdot 10^{-3} \cdot R$	Direct or substitution method.

Original Date of Accreditation: 1980

Page 2 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%

Accreditation Manager

ANNEXURE A

Accreditation No.: 106
Date of Issue: 31 January 2022
Expiry Date: 30 January 2027

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT AND RANGE OF MEASURED QUANTITY	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE
3	DC Current (up to 100 A)			
3.1	DC Current Sources			
3.1.1	Low values (≤ 0,1 mA)			
	1 µA to 100 µA	DC	5•10 ⁻⁵ •I	Direct or current Volt drop method.
3.1.2	Intermediate values (> 0,1 mA to 20 A)			
	100 mA to 2 A	DC	2•10 ⁻⁴ •I	Direct or current Volt drop method.
	2 A to 10 A	DC	2•10 ⁻⁴ •I	
	10 A to 20 A	DC	5•10 ⁻⁴ •I	
3.1.3	High values (> 20 A to 100 A)			
	20 A to 50 A	DC	5•10 ⁻⁴ •I	Direct or current Volt drop method.
	50 A to 100 A	DC	2•10 ⁻³ •I	
3.2	Current Meters			
3.2.1	Low values (<0,1 mA)			
	1 µA to 100 µA	DC	5•10 ⁻⁵ •I	Direct measurement.
3.2.2	Intermediate values (0,1 mA to 20 A)			
	0,1 mA to 100 mA	DC	2•10 ⁻⁴ •I	Direct measurement.
	100 mA to 2 A	DC	2•10 ⁻⁴ •I	
	2 A to 10 A	DC	2•10 ⁻⁴ •I	
	10 A to 20 A	DC	2•10 ⁻³ •I	
3.2.3	High values (>20 A to 100 A)			
	20 A to 50 A	DC	2•10 ⁻³ •I	Direct measurement.
4	Impedence (up to the MHz range)			
4.2	Capacitance			
4.2.1	Standard Capacitors			
	100 pF to 1 µF	120 Hz to 10 kHz	1•10 ⁻³ •C	Direct measurement or substitution method.
	1 µF to 10 µF	120 Hz to 10 kHz	2•10 ⁻³ •C	
	10 µF to 100 µF	120 Hz to 1 kHz	2•10 ⁻³ •C	
4.2.2	Fixed Capacitor			
	100 pF to 1 µF	120 Hz to 10 kHz	1•10 ⁻³ •C	Direct measurement or substitution method.
	1 µF to 10 µF	120 Hz to 10 kHz	2•10 ⁻³ •C	
	10 µF to 100 µF	120 Hz to 1 kHz	2•10 ⁻³ •C	
4.2.4	Capacitance Meters and Bridges			
	1 pF, 10 pF, 100 pF and 1 nF	50 Hz to 1 kHz 1 kHz to 1 MHz	5•10 ⁻⁵ •C 5•10 ⁻⁴ •C	Direct measurement of reference standard capacitor.
	10 nF	50 Hz to 1 kHz	5•10 ⁻⁵ •C	
		1 kHz to 100 kHz	5•10 ⁻⁴ •C	
		100 kHz to 500 kHz 500 kHz to 1 MHz	1•10 ⁻³ •C 1•10 ⁻² •C	
	100 nF	50 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 200 kHz	5•10 ⁻⁵ •C 5•10 ⁻⁴ •C 3•10 ⁻³ •C	
	1 µF	50 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 50 kHz	5•10 ⁻⁵ •C 1•10 ⁻³ •C 2•10 ⁻³ •C	
	10 µF and 100 µF	120 Hz to 1 kHz	2•10 ⁻³ •C	

Original Date of Accreditation: 1980

Page 3 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%

Accreditation Manager

ANNEXURE A

Accreditation No.: 106
Date of Issue: 31 January 2022
Expiry Date: 30 January 2027

ITEM	MEASURED QUANTITY, TYPE OF GAUGE OR INSTRUMENT AND RANGE OF MEASURED QUANTITY	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE
4	Impedance (up to the MHz range)			
4.3	Inductance			
4.3.2	Fixed Indicator, Variable, Indicator box (1 mH to 1 H)			
	1 mH to 1 H	120 Hz to 1 kHz	2•10 ⁻³ •L	Direct measurement or substitution method.
4.3.3	Fixed Indicator, Variable, Indicator box (> 1 H)			
	1 H to 10 H	120 Hz	5•10 ⁻³ •L	Direct measurement or substitution method.
4.3.5	Inductance Meters and Bridges			
	1 mH to 1 H	120 Hz to 1 kHz	2•10 ⁻³ •L	Direct measurement.
	1 H to 10 H	120 Hz	5•10 ⁻³ •L	
5	AC Voltage			
5.2.1	AC Voltage Sources (up to 1 000 V)			
	0 V to 10 V	30 Hz to 20 kHz 20 kHz to 100 kHz 100 kHz to 1 MHz	1•10 ⁻⁴ •U + 10 μV 2•10 ⁻⁴ •U + 10 μV 3•10 ⁻⁴ •U + 10 μV	Direct measurement.
	10 V to 20 V	30 Hz to 20 kHz 20 kHz to 100 kHz	1•10 ⁻⁴ •U 2•10 ⁻⁴ •U	
	20 V to 200 V	30 Hz to 20 kHz 20 kHz to 100 kHz	1•10 ⁻⁴ •U 2•10 ⁻⁴ •U	
	200 V to 1 000 V	30 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 100 kHz	3•10 ⁻⁴ •U 4•10 ⁻⁴ •U 6•10 ⁻⁴ •U	
5.2	AC Voltage (up to 1 000 V)			
5.2.2	Meters			
	0 V to 10 V	30 Hz to 20 kHz 20 kHz to 100 kHz 100 kHz to 1 MHz	1•10 ⁻⁴ •U + 10 μV 2•10 ⁻⁴ •U + 10 μV 3•10 ⁻⁴ •U + 10 μV	Direct measurement.
	10 V to 20 V	30 Hz to 20 kHz 20 kHz to 100 kHz 100 kHz to 500 kHz 500 kHz to 1 MHz	1•10 ⁻⁴ •U 2•10 ⁻⁴ •U 3•10 ⁻⁴ •U 4•10 ⁻⁴ •U	
	20 V to 200 V	30 Hz to 20 kHz 20 kHz to 100 kHz	1•10 ⁻⁴ •U 2•10 ⁻⁴ •U	
	200 V to 1 000 V	50 Hz to 1 kHz	3•10 ⁻⁴ •U	
6	AC current			
6.2	AC current (up to 100 A)			
6.2.1	Sources			
	0 A to 2 A	30 Hz to 5 kHz	3•10 ⁻⁴ •I + 2 μA	Direct measurement or volt drop method.
6.2.2	Meter			
	0 A to 2 A	30 Hz to 5 kHz	3•10 ⁻⁴ •I + 2 μA	Direct measurement or volt drop method.

Original date of accreditation: 1980

Page 4 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%

ISSUED BY THE SOUTH AFRICAN NATIONAL ACCREDITATION SYSTEM

Accreditation Manager