# **CERTIFICATE OF ACCREDITATION**

In terms of section 22(2) (b) of the Accreditation for Conformity Assessment, Calibration and Good Laboratory Practice Act, 2006 (Act 19 of 2006), read with sections 23(1), (2) and (3) of the said Act, I hereby certify that:-

# NATIONAL METROLOGY INSTITUTE OF SOUTH AFRICA ELECTRICAL DCLF LABORATORY

Accreditation Number: 1612

is a South African National Accreditation System accredited Calibration Laboratory provided that all SANAS conditions and requirements are complied with

This certificate is valid as per the scope as stated in the accompanying scope of accreditation, Annexure "A", bearing the above accreditation number for

# DC LOW FREQUENCY METROLOGY

The facility is accredited in accordance with the recognised International Standard

# ISO/IEC 17025:2017

The accreditation demonstrates technical competency for a defined scope and the operation of a laboratory quality management system

While this certificate remains valid, the Accredited Facility named above is authorised to use the relevant SANAS accreditation symbol to issue facility reports and/or certificates

Mr T Baleni Acting Chief Executive Officer

Effective Date: 02 August 2021 Certificate Expires: 10 July 2026

# SCOPE OF ACCREDITATION

DC LOW FREQUENCY METROLOGY

Accreditation Number: 1612

Permanent Address of Laboratory: National Metrology Institute of SA Electrical DCLF Laboratory Building 5, CSIR Campus Meiring Naude Road Brummeria Pretoria 0001		<u>Technical Signat</u>	ories: Mr AM Matlejoane (It Mr M Khoza (Items 1. Mr PJ Prinsloo (Items Dr E Golovins (Item 5 Mr M Hlakola (Items	ems 1, 2, 3 & 4) 1.1b, 1.1.2, 1.1.3, 1.2, 2, 3 & 4) 5, 6 & 7) 5.1, 5.2, 6.1 & 6.2) 1.1.1b, 1.1.2, 1.1.3,1.2, 2.1, 2.3, 4, 4.2 & 4.3.5)	
<u>Postal Address:</u> Private Bag X 34 Lynnwood Ridge 0040		Nominated Repre	esentative: Ms LN Ntatamala		
Tel: (012) 947-2781 Email: Intatamala@nmisa.org		Issue No.: Date of Issue: Expiry Date:	16 26 April 2022 10 July 2026		
ITEM	NOMINAL RANGE	F	NOMINAL REQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE
1	DC Voltage				
1.1	DC Voltage sources				
1.1.1	Single values (≤ 10 V) DC	referenc	e standards (Josepl	hson Voltage)	
(a)	1 V 1,018 V 10 V		DC DC DC	1∙10 <sup>-7</sup> •U 1•10 <sup>-7</sup> •U 2•10 <sup>-8</sup> •U	Direct comparison with Josephson voltage standard.
1.1.1	Single values (≤ 10 V) DC reference standards (8 000A)				
(b)	1 V 1,018 V 10 V		DC DC DC	4•10 <sup>-7</sup> •U 4•10 <sup>-7</sup> •U 2•10 <sup>-7</sup> •U	Direct comparison with DC voltage reference standard using an automatic potentiometer.
1.1.2	Low values (≤ 10 V) Multifunction calibrator				
	100 mV 1 V 10 V		DC DC DC	12•10 <sup>-6</sup> •U 1,6•10 <sup>-6</sup> •U 1,2•10 <sup>-6</sup> •U	Direct comparison with DC voltage reference standard using a reference divider.

Original date of accreditation: 01 September 2005

Page 1 of 9

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%

#### Accreditation No.: 1612 Date of Issue: 26 April 2022 Expiry Date: 10 July 2026

ITEM	NOMINAL RANGE	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE	
1.1.3	Intermediate values (> 10 V	to 1100 V) Multifunction calib	orator		
	100 V 1 000 V	DC DC	1,5•10 <sup>-6</sup> •U 1•10 <sup>-5</sup> •U	Direct comparison with DC voltage reference standard using a reference divider.	
1.2	DC Voltage meters			·	
1.2.1	Very low values (≤ 1 mV) Dig	gital multimeter, nano voltmet	ters		
	0 mV 1 mV		10 nV 5•10⁻⁵•U	Direct comparison with DC voltage source using a resistive divider and short.	
1.2.2	Intermediate values (> 1 mV	to 1100 V) Digital multimeter	rs, nano voltmeters		
	1 mV to 10 mV 10 mV to 100 mV 0,1 V to 1 000 V		5•10 <sup>-5</sup> •U 2•10 <sup>-5</sup> •U 1•10 <sup>-5</sup> •U	Direct comparison with DC voltage source using a resistive divider and direct comparison with DC voltage source.	
2	DC Resistance				
2.1	DC Resistance standards and sources.				
2.1.1	Low values ( $\leq 1 \Omega$ ) Fixed resistor, resistance box, multifunction calibrator				
	10 μΩ to 1 mΩ 1 mΩ to 10 mΩ 10 mΩ to 100 mΩ 100 mΩ to 1 Ω		7•10 <sup>-6</sup> •R 2•10 <sup>-6</sup> •R 2•10 <sup>-6</sup> •R 1•10 <sup>-6</sup> •R	Comparison with reference standard resistors using an automated resistance bridge.	
2.1.2	Intermediate values (> 1 Ω to	o 1 M $\Omega$ ) Fixed resistor, resist	ance box, multifunction calib	rator	
	1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ		2•10 <sup>-7</sup> •R 3•10 <sup>-7</sup> •R 4•10 <sup>-7</sup> •R 5•10 <sup>-7</sup> •R 2•10 <sup>-6</sup> •R 4•10 <sup>-6</sup> •R	Comparison with reference standard resistors using an automated resistance bridge.	

Original date of accreditation: 01 September 2005

Page 2 of 9

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%

#### Accreditation No.: 1612 Date of Issue: 26 April 2022 Expiry Date: 10 July 2026

ITEM	NOMINAL RANGE	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PRECEDURE
2.1.3	High values (> 1 M $\Omega$ ) Fixed r	esistor, resistance box, multi	function calibrator	
	1 MΩ to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 1 GΩ 1 GΩ to 10 GΩ 10 GΩ to 100 GΩ 100 GΩ to 1 TΩ		6•10 <sup>-6</sup> •R 6•10 <sup>-6</sup> •R 9•10 <sup>-6</sup> •R 7•10 <sup>-5</sup> •R 3•10 <sup>-4</sup> •R 2•10 <sup>-3</sup> •R	Comparison with reference standard resistors using an automated resistance bridge.
2.2	DC resistance meters			
2.2.1	Low values (≤ 1 Ω) Micro-ohi	m meter, multimeter, resistar	nce bridge	
	0 Ω 1 Ω		0,3 μΩ 1•10 <sup>-6</sup> •R	Direct comparison with standard resistor or short.
2.2.2	Intermediate values (> 1 $\Omega$ to	1 G $\Omega$ ) ohm meter, multimet	er, resistance bridge	
	10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ		2 •10 <sup>-6</sup> •R 3•10 <sup>-6</sup> •R 3•10 <sup>-6</sup> •R	
	10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ		3•10 <sup>-₅</sup> •R 5•10 <sup>-6</sup> •R 7•10 <sup>-6</sup> •R 9•10 <sup>-6</sup> •R	Direct comparison with standard resistors.
	100 MΩ to 1 GΩ		2•10 <sup>-5</sup> •R	
2.2.3	High values (> 1 G $\Omega$ ) multime	eter, multifunctional transfer	std teraohmmeter	
	10 GΩ to 100 GΩ 100 GΩ to 1 TΩ		7•10 <sup>-3</sup> •R 3•10 <sup>-4</sup> •R 2•10 <sup>-3</sup> •R	Direct comparison with standard resistors.
3	DC current (up to 100A)			
3.1	DC current sources			
3.1.1	Low values (≤ 0,1 mA) Current generator, multifunction calibrator.			
	0,1 μA to 1 μA 1 μA to 10 μA 10 μA to 100 μA	DC DC DC	2•10 <sup>-4</sup> •l 5•10 <sup>-5</sup> •l 3•10 <sup>-5</sup> •l	Current volt-drop method using standard resistors and DMM.
3.1.2	Intermediate values (> 0,1 m	A to 20 A) Current generator	, multifunction calibrator	
	0,1 mA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 0,1 A to 1 A	DC DC DC DC	3•10 <sup>-5</sup> •I 3•10 <sup>-5</sup> •I 3•10 <sup>-5</sup> •I 3•10 <sup>-5</sup> •I	Current volt-drop method using standard resistors and DMM.
	1 A to 20 A	DC	3•10 <sup>-5</sup> •I	

Original date of accreditation: 01 September 2005

Page 3 of 9

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%

#### Accreditation No.: 1612 Date of Issue: 26 April 2022 Expiry Date: 10 July 2026

ITEM	NOMINAL RANGE	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE
3.1.3	High values (> 20 A to 100 A	) Current generator.		
	20 A to 100 A	DC	<b>2•10</b> <sup>-4</sup> •l	Current volt-drop method using standard resistors and DMM.
3.2	DC current meters.			
3.2.1	Low values (≤ 0,1 mA) pico a	mmeter, nano ammeter, multi	meter.	
	0 μΑ 0,1 μΑ to 1 μΑ 1 μΑ to 10 μΑ 10 μΑ to 100 μΑ		0,04 µA 2•10 <sup>-4</sup> •l 5•10 <sup>-5</sup> •l 3•10 <sup>-5</sup> •l	Direct comparison with current source output.
3.2.2	Intermediate values (> 0,1 m	A to 20 A) multimeter, current	meter	
	0,1 mA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 0,1 A to 1 A 1 A to 20 A		3•10 <sup>-5</sup> •l 3•10 <sup>-5</sup> •l 3•10 <sup>-5</sup> •l 3•10 <sup>-5</sup> •l 3•10 <sup>-5</sup> •l 3•10 <sup>-5</sup> •l	Direct comparison with current source output.
4	Impedance (up to 200 kH	lz range)		
4.2	Capacitance			
4.2.1	Capacitance and dissipation	factor for low loss capacitors,	standard capacitor (sealed, dry-	nitrogen, fused silica)
	1 pF to 100 pF 1 nF to 100 nF 1 μF 10 μF 1 pF to 1 μF 10 μF 1 pF to 100 nF 1 μF 1 pF 10 pF to 10 nF 100 nF	1 kHz 1 kHz 1 kHz 1 kHz 10 kHz 10 kHz 100 kHz 100 kHz 1 MHz 1 MHz 1 MHz 1 MHz	$4 \cdot 10^{-7} \cdot C$ $6 \cdot 10^{-6} \cdot C$ $5 \cdot 10^{-5} \cdot C$ $2 \cdot 10^{-4} \cdot C$ $4 \cdot 10^{-4} \cdot C$ $6 \cdot 10^{-4} \cdot C$ $3 \cdot 10^{-4} \cdot C$ $8 \cdot 10^{-4} \cdot C$ $2 \cdot 10^{-3} \cdot C$ $8 \cdot 10^{-4} \cdot C$ $2 \cdot 10^{-3} \cdot C$	Comparison with reference standard capacitors using an LCR bridge.

Original date of accreditation: 01 September 2005

Page 4 of 9

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%

#### Accreditation No.: 1612 Date of Issue: 26 April 2022 Expiry Date: 10 July 2026

ITEM	NOMINAL RANGE	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE
4.2.2	Capacitance and dissipation f	actor for dielectric capaci	tors, fixed capacitor, switched cap	pacitor, capacitance box
	1 pF to 100 pF	1 kHz	4•10 <sup>-7</sup> •C	
	1 nF to 100 nF	1 kHz	6•10 <sup>-6</sup> •C	
	1μF	1 kHz	5•10 <sup>-5</sup> •C	
	10 μF	1 kHz	2•10 <sup>-4</sup> •C	
	1 pF to 1 μF	10 kHz	4•10 <sup>-4</sup> •C	Comparison with reference
	10 μF	10 kHz	6•10 <sup>-4</sup> •C	standard capacitors using
	1 pF to 100 nF	100 kHz	3•10 <sup>-4</sup> •C	a LCR bridge.
	1 μF	100 kHz	8•10 <sup>-4</sup> •C	
	1 pF	1 MHz	2•10 <sup>-3</sup> •C	
	10 pF to 10 nF	1 MHZ	8•10 <sup>-4</sup> •C	
	100 nF	I MHZ	2•10 ••0	
4.2.3	Capacitance and dissipation f	actor for transformed cap	acitors, fixed capacitor, switched	capacitor
	1 pF to 100 pF	1 kHz	4•10 <sup>-7</sup> •C	
	1 nF to 100 nF	1 kHz	6•10 <sup>-6</sup> •C	
	1 μF	1 kHz	5•10 <sup>-</sup> •C	
	10 μF	1 kHz	2•10 <sup>-4</sup> •C	Comparison with reference
	1 pF to 1 μF	10 kHz	4•10 <sup>-4</sup> •C	standard capacitors using
	10 μF	10 kHz	6•10 <sup></sup> •C	a LCR bridge.
	1 pF to 100 nF	100 kHz	3•10 <sup>-</sup> •C	5
	1 µF	100 KHZ	2•10 <sup>-3</sup> •C	
			8•10 <sup>-4</sup> •C	
			2•10 <sup>-3</sup> •C	
			210 0	
4.2.4	Meters, capacitance bridge, L	CR meter		•
	1 pF to 100 pF	1 kHz	4•10 <sup>-7</sup> •C	
		1 KHZ	3•10 <sup>-4</sup> •C	
			7•10 <sup>-3</sup> •C	
	$10 \mu F$	1 KHZ	3•10 ••0	
		10 KHZ 10 kHz	6•10 <sup>-4</sup> •C	Comparison with reference
	1 pE to 100 pE		4•10 <sup>-4</sup> •C	standard capacitors using
	1 uF	100 kHz	8•10 <sup>-4</sup> •C	a LON DHuye.
	1 pF	1 MHz	2•10 <sup>-3</sup> •C	
	1 pF to 10 nF	1 MHz	- 10 ℃ 8•10 <sup>-3</sup> •C	
	100 nF	1 MHz	2•10 <sup>-3</sup> •C	

Original date of accreditation: 01 September 2005

Page 5 of 9

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%

Accreditation No.: 1612 Date of Issue: 26 April 2022 Expiry Date: 10 July 2026

ITEM	NOMINAL RANGE	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE		
4.3	Inductance					
4.3.1	Low values (< 1 mH) self-inductance and equivalent series resistance, fixed inductor, variable inductor, inductance box					
	100 μH	1 kHz	3•10⁻⁴•L	Comparison with reference standard inductors using an LCR bridge		
4.3.2	Intermediate values (≥1 n inductance box	nH to 1 H) self-inductance ar	nd equivalent series resistance	e, fixed inductor, variable inductor,		
	1 mH	1 kHz	7•10 <sup>-5</sup> •L			
	10 mH	1 kHz	7•10 <sup>-5</sup> •L	Comparison with reference		
	100 mH	1 kHz	7•10 <sup>-5</sup> •L	standard inductors using an LCR		
	1 H	1 Hz	7•10⁻⁵•L	bridge		
	1 H	100 Hz	8•10⁻⁵•L			
4.3.3	High values (> 1 H) self ir	nductance and equivalent se	ries resistance, fixed inductor,	variable inductor, inductance box		
	10 H 10 H	100 Hz 1 kHz	8•10 <sup>-5</sup> •L 4•10 <sup>-5</sup> •L	Comparison with reference standard inductors using an LCR bridge.		
4.3.5	Meters, LCR meter	Meters, LCR meter				
	100 μH	1 kHz	3•10⁻⁴•L			
	1 mH	1 kHz	7•10 <sup>-5</sup> •L			
	10 mH	1 kHz	7•10 <sup>-5</sup> •L			
	100 mH	1 kHz	7•10⁻⁵•L	Direct comparison with standard		
	1 H	1 kHz	7•10⁻⁵•L	inductors.		
	10 H	1 kHz	4•10 <sup>-5</sup> •L			
	1 H	100 Hz	8•10 <sup>-5</sup> •L			
	10 H	100 Hz	8•10 <sup>-3</sup> •L			
5	AC voltage (up to the	AC voltage (up to the MHz range)				
5.1	AC/DC voltage transfer					
5.1.1	AC/DC transfer difference	e at low voltages (typically be	low or equal to 500 mV)			
	2 mV to 200 mV 2 mV to 200 mV 200 mV to 0,5 mV 0,2 V to 0,5 V	10 Hz to 300 kHz 300 kHz to 1 MHz 10 Hz to 1 MHz 10 Hz to 1 MHz	400 μV/V 900 μV/V 60 μV/V 60 μV/V	Micro-potentiometer step-down measurement; Direct comparison with reference standard		

Original date of accreditation: 01 September 2005

Page 6 of 9

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%

#### Accreditation No.: 1612 Date of Issue: 26 April 2022 Expiry Date: 10 July 2026

ITEM	NOMINAL RANGE	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE
5.1.2	AC/DC transfer difference at low voltages (typically above 500 mV to 5 V)			
	0,5 V to 5 V	10 Hz to 1 MHz	60 μV/V	Direct comparison with reference standard
5.1.3	AC/DC transfer difference at hi	igher voltages (typically 5 \	/)	
	5 V to 30 V 5 V to 30 V 30 V to 1 000 V	10 Hz to 100 kHz 100 kHz to 1 MHz 10 Hz to 100 kHz	30 μV/V 100 μV/V 70 μV/V	Direct comparison with reference standard
5.2	AC voltage up to 1100 V			
5.2.1	Sources, multifunction calibrate	or		
	2 mV to 30 mV 30 mV to 200 mV 200 mV to 30 V 200 mV to 20 V 30 V to 1 000 V	10 Hz to 1 MHz 10 Hz to 1 MHz 10 Hz to 100 kHz 100 kHz to 1 MHz 10 Hz to 100 kHz	1,3•10 <sup>-3</sup> •U 3•10 <sup>-4</sup> •U 4•10 <sup>-5</sup> •U 1•10 <sup>-4</sup> •U 1•10 <sup>-4</sup> •U	Measurement with AC/DC transfer standard
5.2.2	Meters, ac voltmeter			
	2 mV to 200 mV 2 mV to 200 mV 200 mV to 30 V 200 mV to 30 V 30 V to 1 000 V	10 Hz to 300 kHz 300 kHz to 1 MHz 10 Hz to 100 kHz 100 kHz to 1 MHz 10 Hz to 100 kHz	5•10 <sup>-4</sup> •U 1•10 <sup>-3</sup> •U 5•10 <sup>-5</sup> •U 1•10 <sup>-4</sup> •U 7•10 <sup>-5</sup> •U	Measurement with AC/DC transfer standard
6	AC current			
6.1	AC/DC current transfer			
6.1.1	AC/DC transfer difference			
	2 mA to 5 mA 2 mA to 5 mA 2 mA to 5 mA 5 mA to 20 mA 5 mA to 20 mA 5 mA to 20 mA 20 mA to 20 mA 20 mA to 50 mA 20 mA to 50 mA 20 mA to 50 mA 50 mA to 300 mA 50 mA to 300 mA 50 mA to 300 mA	10 Hz to 5 kHz 5 kHz to 50 kHz 50 kHz to 100 kHz 10 Hz to 5 kHz 5 kHz to 50 kHz 50 kHz to 100 kHz 10 Hz to 5 kHz 5 kHz to 50 kHz 50 kHz to 100 kHz 10 Hz to 20 Hz 20 Hz to 5 kHz 5 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 50 kHz	20 μA/A 25 μA/A 41 μA/A 16 μA/A 20 μA/A 20 μA/A 20 μA/A 27 μA/A 27 μA/A 23 μA/A 28 μA/A 37 μA/A 51 μA/A	Direct comparison with reference standard.

Original date of accreditation: 01 September 2005

Page 7 of 9

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%

#### Accreditation No.: 1612 Date of Issue: 26 April 2022 Expiry Date: 10 July 2026

ITEM	NOMINAL RANGE	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE
6.1.1	AC/DC transfer difference (	Continued from previous p	age)	
	300 mA to 2 A 300 mA to 2 A 300 mA to 2 A 300 mA to 1 A 300 mA to 1 A 1 A to 3 A 2 A to 5 A 2 A to 5 A 2 A to 5 A 2 A to 100 A 3 A to 100 A 5 A to 100 A 5 A to 100 A 10 A to 100 A	10 Hz to 20 Hz 20 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 50 kHz 20 kHz to 50 kHz 50 kHz to 50 kHz 10 Hz to 20 Hz 20 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 10 Hz to 20 Hz 20 Hz to 10 kHz 50 kHz to 100 kHz	41 μA/A 29 μA/A 42 μA/A 47 μA/A 73 μA/A 76 μA/A 140 μA/A 65 μA/A 34 μA/A 73 μA/A 120 μA/A 210 μA/A 76 μA/A 40 μA/A 240 μA/A	Direct comparison with reference standard
6.2	AC current up to 100A			
6.2.1	Sources, Multifunction calib	orator		
	10 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 11 A	20 Hz to 5 kHz 20 Hz to 5 kHz 20 Hz to 5 kHz 20 Hz to 5 kHz	6•10 <sup>-5</sup> •l 9•10 <sup>-5</sup> •l 2•10 <sup>-4</sup> •l 6•10 <sup>-4</sup> •l	Measurement with AC/DC transfer standard and shunt
6.2.2	Meters, ac meter, multimete	er, multifunction transfer sta	andard	
	10 mA to 200 mA 200 mA to 2 A 2 A to 11 A 2 A to 11 A	20 Hz to 5 kHz 20 Hz to 5 kHz 20 Hz to 1 kHz 1 kHz to 5 kHz	5•10 <sup>-4</sup> •l 9•10 <sup>-4</sup> •l 6•10 <sup>-4</sup> •l 4•10 <sup>-3</sup> •l	Measurement with AC/DC transfer standard and shunt; Direct comparison with reference source.

Original date of accreditation: 01 September 2005

Page 8 of 9

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%

ISSUED BY THE SOUTH AFRICAN NATIONAL ACCREDITATION SYSTEM

#### Accreditation No.: 1612 Date of Issue: 26 April 2022 Expiry Date: 10 July 2026

ITEM	NOMINAL RANGE	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE
7	AC power			
7.1	AC power and energy			
7.1.1	Single phase (frequencies I	below or equal to 400Hz)		
	Active/ Reactive Power and Energy 30 V to 480 V 10 mA to 50 mA Power factor 0 to 1 inductive and capacitive	48 Hz to 63 Hz	3•10⁻⁴•VA	Direct comparison with reference standard
	Active/ Reactive Power and Energy 30 V to 480 V 50 mA to 160 A Power factor 0 to 1 inductive and capacitive	48 Hz to 63 Hz	8•10 <sup>-5</sup> •VA	Direct comparison with reference standard

Original date of accreditation: 01 September 2005

Page 9 of 9

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%

ISSUED BY THE SOUTH AFRICAN NATIONAL ACCREDITATION SYSTEM