

# **CERTIFICATE OF ACCREDITATION**

*In terms of section 22(2) (b) of the Accreditation for Conformity Assessment, Calibration and Good Laboratory Practice Act, 2006 (Act 19 of 2006), read with sections 23(1), (2) and (3) of the said Act, I hereby certify that:-*

**LMC SOLUTIONS (PTY) LTD**  
**Co. Reg. No.: 2015/145753/07**  
**TEMPERATURE CALIBRATION LABORATORY**

Accreditation Number: **375**

is a South African National Accreditation System accredited Calibration laboratory provided that all SANAS conditions and requirements are complied with

This certificate is valid as per the scope as stated in the accompanying scope of accreditation Annexure "A", bearing the above accreditation number for

## **TEMPERATURE METROLOGY**

The facility is accredited in accordance with the recognised International Standard

**ISO/IEC 17025:2017**

The accreditation demonstrates technical competency for a defined scope and the operation of a laboratory quality management system

While this certificate remains valid, the Accredited Facility named above is authorised to use the relevant SANAS accreditation symbol to issue facility reports and/or certificates

---

**Mr T Baleni**  
**Acting Chief Executive Officer**

**Effective Date: 10 October 2021**  
**Certificate Expires: 09 October 2026**

## ANNEXURE A

## SCOPE OF ACCREDITATION TEMPERATURE METROLOGY

Accreditation Number 375

<b>Permanent Address of Laboratory:</b> LMC Solutions (Pty) Ltd Temperature Calibration Laboratory 56 Rubida Street Lynnwood Ridge Pretoria 0081		<b>Technical Signatories:</b> Mr M Sakharov (Item 1) Mr H Liedberg		
<b>Postal Address:</b> 56 Rubida Street Lynnwood Ridge Pretoria 0081		<b>Nominated Representative:</b> Mr H Liedberg		
Cell: +2782 740-6557	Fax: n/a	E-mail: <a href="mailto:info@LMC-Solutions.co.za">info@LMC-Solutions.co.za</a>	Issue No.: 05 Date of Issue: 10 October 2021 Expiry Date: 09 October 2026	
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	METHOD/ PROCEDURE
<b>1</b>	<b>THERMOMETRY</b>			
<b>1.1</b>	<b>Thermocouples</b>			
1.1.1	Noble Metal	- 40 °C to 0 °C 0 °C to 250 °C 250 °C to 450 °C 450 °C to 1200 °C	0,75 K + (0,009 *  t ) K 0,75 K - (0,0012 *  t ) K 0,5 K 0,35 K + (0,0013* t ) K	Calibration by comparison with a reference thermometer in a bath, drywell, or furnace.
1.1.2	Base Metal	- 80 °C to 0 °C 0 °C to 250 °C 250 °C to 450 °C 450 °C to 1200 °C	0,11 K + (0,00065 *  t ) K 0,11 K + (0,00025 *  t ) K 0,5 K 0,35 K + (0,0013* t ) K	
<b>1.2</b>	<b>Resistance Thermometers</b>			
1.2.1	Platinum Resistance Thermometers (PT 100)	- 80 °C to 250 °C 250 °C to 450 °C	0,05 K + (0,0004 *  t ) K 0,07 K + (0,0004* t ) K	Calibration by comparison with a reference thermometer in a bath, drywell, or furnace
<b>1.3</b>	<b>Thermometers</b>			
1.3.1	Liquid-in-glass	- 40 °C to 0 °C 0 °C to 250 °C	0,05 K + (0,0005 *  t ) K 0,05 K + (0,00045 *  t ) K	Calibration by comparison with a reference thermometer in a bath, drywell, or furnace

Original Date of Accreditation: 10 October 2016

Page 1 of 2

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor  $k = 2$ , corresponding to a confidence level of approximately 95%

### Accreditation Manager

## ANNEXURE A

Accreditation No.: 375  
Date of Issue: 10 October 2021  
Expiry Date: 09 October 2026

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	METHOD / PROCEDURE
1.3.2	Digital Thermometers	- 80 °C to 250 °C 250 °C to 450 °C 450 °C to 1 200 °C	0,05 K + (0,0004 *  t ) K 0,07 K + (0,0004* t ) K 0,35 K + (0,0013* t ) K	Calibration by comparison with a reference thermometer in a bath, drywell, or furnace.
1.3.5	Radiation Thermometers	0 °C to 20 °C 20 °C to 40 °C 40 °C to 80 °C	1,0 K 0,5 K 0,6 K	Calibration using a radiation source and reference thermometer.
<b>1.4</b>	<b>Reference Temperature Sources</b>			
1.4.1	Ice Point Reference	0,0 °C	0,02 K	Prepared in a thermally insulated flask using distilled water and Ice.
<b>1.5</b>	<b>Temperature Measuring and Recording</b>			
1.5.2	Data Loggers	- 80 °C to 250 °C 250 °C to 450 °C 450 °C to 1 200 °C	0,2 K 0,25 K 0,35 K + (0,0013* t ) K	Calibration by comparison with a reference thermometer in a chamber.
<b>4</b>	<b>TEMPERATURE INSTALLATIONS AND DEVICES</b>			
<b>4.1</b>	<b>Iso-thermal Media evaluation (multi point over time monitoring)</b>			
4.1.2	Environmental Chambers	- 80 °C to 121 °C	0,2 K	Calibration by temperature mapping over time using reference thermometers and/or loggers.
4.1.3	Furnaces / Drying Ovens	121 °C to 150 °C	0,22 K	
4.1.4	Fridges / Freezers	150 °C to 450 °C	1,0 K	
4.1.5	Incubators	450 °C to 1 200 °C	0,4 K + (0,0015* t ) K	
4.1.6	Liquid Baths			
<b>5</b>	On-site calibration for items 1 and 4 above			

Original Date of Accreditation: 10 October 2016

Page 2 of 2

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor  $k = 2$ , corresponding to a confidence level of approximately 95%

ISSUED BY THE SOUTH AFRICAN NATIONAL ACCREDITATION SYSTEM

**Accreditation Manager**