CERTIFICATE OF ACCREDITATION

In terms of section 22(2) (b) of the Accreditation for Conformity Assessment, Calibration and Good Laboratory Practice Act, 2006 (Act 19 of 2006), read with sections 23(1), (2) and (3) of the said Act, I hereby certify that:-

CONCILIUM TECHNOLOGIES (PTY) LTD Co. Reg. No.: 1999/013330/07 RADIO FREQUENCY CALIBRATION LABORATORY

Accreditation Number: 706

is a South African National Accreditation System accredited Calibration laboratory provided that all SANAS conditions and requirements are complied with

This certificate is valid as per the scope as stated in the accompanying scope of accreditation Annexure "A", bearing the above accreditation number for

RADIO FREQUENCY METROLOGY

The facility is accredited in accordance with the recognised International Standard

ISO/IEC 17025:2017

The accreditation demonstrates technical competency for a defined scope and the operation of a laboratory quality management system

While this certificate remains valid, the Accredited Facility named above is authorised to use the relevant SANAS accreditation symbol to issue facility reports and/or certificates

Mr T Baleni Acting Chief Executive Officer Effective Date: 31 January 2022 Certificate Expires: 30 January 2027

SCOPE OF ACCREDITATION

RADIO FREQUENCY METROLOGY

Accreditation Number: 706

Permanent Address of Laboratory:		Technical Signatories	s: Mr BJH Bremm	Mr BJH Bremmer	
Concilium Technologies (Pty) Ltd			Mr GD Schuste	r	
1 Standford Office Park					
12 Bauhinia Street					
Highveld	d Technopark				
Centurio	n				
0157					
		Nominated Panroson	tativa Mr. P. I.U. Promm	or	
Postal Address:		Nominated hepresen		ei	
P O Box	.6/611				
Highveic	1				
0169					
Tel:	(012) 678-9200	Issue No.:	21		
Fax:	(012) 665-4160	Date of Issue: 31 January 2022		2	
E-mail:	bart bremmer@concilium.co.za	Expiry Date:	30 January 2027		
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT AND NOMINAL RANGE	NOMINAL FREQUENCY /	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN	METHOD / PROCEDURE	
- 1 - 1		<u> </u>			
•••	RADIO FREQUENCY MEASUREMENTS				
11.1	Radio Frequency Power				
11.1.1	Absolute power in coaxial line: power meter, power source				
	0 dBm (1 mW)	50 MHz	0,08 dB		
i)	-60 dBm to -20 dBm (1 nW to 10 μW) -30 dBm to +20 dBm (1 μW to 100 mW) -60 dBm to +20 dBm (1 nW to 100 mW) -60 dBm to +20 dBm (1 nW to 100 mW)	10 MHz to 1 GHz 100 kHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 18 GHz	0,2 dB 0,2 dB 0,3 dB 0,4 dB	Direct measurement method using a power source or comparative measurement using a reference standard or	
	-110 dBm to -30 dBm (10 fW to 1 μW) -110 dBm to - 60 dBm (10 fW to 1 nW) -110 dBm to - 60 dBm (10 fW to 1 nW) N-Type Connector (50 Ω)	100 kHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 18 GHz	0,2 dB + 0,005 dB per dBm 0,3 dB + 0,010 dB per dBm 0,4 dB + 0,015 dB per dBm	indirect method through a directional couple or splinter.	

Original Date of Accreditation: 1980

* Generation Limitation: 100 kHz to 26,5 GHz: + 13 dBm ** Generation Limitation: 26,5 GHz to 50 GHz: +4 dBm

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%

Accreditation Manager

Page 1 of 4

Accreditation No.: 706 Date of Issue: 31 January 2022 Expiry Date: 30 January 2027

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT AND NOMINAL RANGE	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE
ii)	-30 dBm to +20 dBm (1 μW to 100 mW) -30 dBm to +20 dBm (1 μW to 100 mW) -30 dBm to +20 dBm (1 μW to 100 mW)	50 MHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 26,5 GHz	0,2 dB 0,3 dB 0,4 dB	
")	-100 dBm to -30 dBm (10 fW to 1 μW) -100 dBm to -60 dBm (10 fW to 1 nW) -100 dBm to -60 dBm (10 fW to 1 nW) PC – 3.5 Connector (50 Ω)	10 MHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 26,5 GHz	0,2 dB + 0,005 dB per dBm 0,3 dB + 0,010 dB per dBm 0,4 dB + 0,015 dB per dBm	
	-60 dBm to - 20 dBm (1 nW to 10 μW) -60 dBm to - 20 dBm (1 nW to 10 μW) -60 dBm to - 20 dBm (1 nW to 10 μW) -60 dBm to - 20 dBm (1 nW to 10 μW)	50 MHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 40 GHz 40 GHz to 50 GHz	0,2 dB 0,3 dB 0,4 dB 0,6 dB	Direct measurement method using a power sensor or comparative
iii)	-30 dBm to + 20 dBm (1 μW to 100 mW) -30 dBm to + 20 dBm (1 μW to 100 mW) -30 dBm to + 20 dBm (1 μW to 100 mW) -30 dBm to + 20 dBm (1 μW to 100 mW)	50 MHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 40 GHz 40 GHz to 50 GHz	0,2 dB 0,3 dB 0,4 dB 0,5 dB	measurement using a reference standard or indirect method through a directional couple or splinter.
	-100 dBm to - 30 dBm (10 fW to 1 μW) -100 dBm to - 30 dBm (10 fW to 1 μW) -100 dBm to - 30 dBm (10 fW to 1 μW) -100 dBm to - 30 dBm (10 fW to 1 μW) PC – 2.4 Connector (50 Ω)	10 MHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 40 GHz 40 GHz to 50 GHz	0,2 dB + 0,005 dB per dBm 0,3 dB + 0,010 dB per dBm 0,4 dB + 0,015 dB per dBm 0,5 dB + 0,020 dB per dBm	
iv)	-30 dBm to +20 dBm (1 μW to 100 mW) (Generation limitation : + 13 dBm)	100 kHz to 2 GHz	0,2 dB	
	-110 dBm to -30 dBm (10 fW to 1 μW) N-Type Connector (75 Ω)	100 kHz to 2 GHz	0,2 dB + 0,005 dB per dBm	
11.2	Scalar RF reflection coefficient and attenuation			
11.2.3	Attenuation in coaxial line: Passive d	evice		
i)	0 dB to 60 dB 60 dB to 80 dB 80 dB to 110 dB	DC DC DC	0,002 dB 0,01 dB 0,15 dB	Voltage ratio method or Substitution method or Direct method.
11.2 11.2.3 i)	-110 dBm to -30 dBm (10 fW to 1 μW) N-Type Connector (75 Ω) Scalar RF reflection coefficient and Attenuation in coaxial line: Passive d 0 dB to 60 dB 60 dB to 80 dB 80 dB to 110 dB DC Attenuation (50 Ω all connector types)	100 kHz to 2 GHz d attenuation evice DC DC DC DC	0,2 dB + 0,005 dB per dBm 0,002 dB 0,01 dB 0,15 dB	Voltage ratio me Substitution me Direct meth

Original Date of Accreditation: 1980

Page 2 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%

Accreditation Manager

Accreditiaton No.:706 Date of Issue: 31 January 2022 Expiry Date: 30 January 2027

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT AND NOMINAL RANGE	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE		
11.2.3	Attenuation in coaxial lin	Attenuation in coaxial line (values in db): Passive device				
	0 dB to 100 dB 0 dB to 100 dB 0 dB to 100 dB	100 kHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 18 GHz	0,04 dB + 0,005 dB per dB 0,1 dB + 0,010 dB per dB 0,15 dB + 0,015 dB per dB			
ii)	100 dB to 120 dB 100 dB to 120 dB 100 dB to 120 dB N-Type Connector (50 Ω)	100 kHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 18 GHz	1 dB 2 dB 3 dB			
iii)	0 dB to 100 dB 0 dB to 100 dB 0 dB to 100 dB 0 dB to 100 dB PC -3.5 Connector (50 Ω)	10 MHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 18 GHz 18 GHz to 26,5 GHz	0,04 dB + 0,005 dB per dB 0,10 dB + 0,010 dB per dB 0,15 dB + 0,015 dB per dB 0,2 dB + 0,015 dB per dB	Power ratio method or Substitution method or Direct method using a RF detector.		
iv)	$\begin{array}{c} 0 \ dB \ to \ 100 \ dB \\ 0 \ dB \ to \ 100 \ dB \\ 0 \ dB \ to \ 100 \ dB \\ 0 \ dB \ to \ 100 \ dB \\ 0 \ dB \ to \ 100 \ dB \\ 0 \ dB \ to \ 100 \ dB \\ PC-2.4 \ Connector \ (50 \ \Omega \) \end{array}$	10 MHz to 1 GHz 1 GHz to 10 GHz 10 GHz to 18 GHz 18 GHz to 26,5 GHz 26,5 GHz to 40 GHz 40 GHz to 50 GHz	0,04 dB + 0,015 dB per dB 0,1 dB + 0,020 dB per dB 0,15 dB + 0,020 dB per dB 0,2 dB + 0,020 dB per dB 0,4 dB + 0,015 dB per dB 0,5 dB + 0,025 dB per dB			
V)	0 dB to 100 dB (75 Ω, N-Type connector)	100 kHz to 2 GHz	0,04 dB + 0,005 dB per dB			
vi)	0 dB to 60 dB 60 dB to 80 dB 80 dB to 110 dB (75 Ω, N-Type connector)	DC DC DC	0,002 dB 0,01 dB 0,15 dB	Voltage ratio method or substitution method or direct measurement method.		
11.2.1	Reflection coefficient in coaxial line: passive device					
i)	Ratio 0 to 1 Ratio 0 to 1 Ratio 0 to 1 N-Type Connector (50 Ω)	5 Hz to 10 MHz 10 MHz to 15 GHz 15 GHz to 18 GHz	$\begin{array}{c} 0,010 + 0,056 \ \rho^2 \\ 0,011 + 0,031 \ \rho^2 \\ 0,013 + 0,100 \ \rho^2 \end{array}$	Calibration against a standard by using directional bridge and/or		
ii)	Ratio 0 to 1 Ratio 0 to 1 Ratio 0 to 1 PC -3.5 Connector (50 Ω)	10 MHz to 15 GHz 15 GHz to 18 GHz 18 GHz to 26,5 GHz	$\begin{array}{c} 0,017 + 0,037 \ \rho^2 \\ 0,013 + 0,100 \ \rho^2 \\ 0,018 + 0,172 \ \rho^2 \end{array}$	ripple extraction method and/or direct measurement method.		

Original Date of Accreditation: 1980

Page 3 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%

Accreditation Manager

Accrediation No.:706 Date of Issue: 31 January 2022 Expiry Date: 30 January 2027

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT AND NOMINAL RANGE	NOMINAL FREQUENCY		CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE
	Ratio 0 to 1 Ratio 0 to 1	10 MHz to 15 GHz 15 GHz to 18 GHz		0,011 + 0,057 ρ ² 0,013 + 0,100 ρ ²	Calibration against a
11.2.1	Ratio 0 to 1	18 GHz to 26,5 GHz		0,018 + 0,172 ρ ²	standard by using
iii)	Ratio 0 to 1	26,5 GHz to 40 GHz		0,050 + 0,167 ρ ²	directional bridge
	Ratio 0 to 1	40 GHz to 50 GHz		0,059 + 0,290 ρ ²	and/or ripple
	PC -2.4 Connector (50 Ω)				extraction method
	Ratio 0 to 1	5 Hz to 10 MHz		0,010 + 0,056 ρ ²	and/or direct measurement
iv)	Ratio 0 to 1 N-Type Connector (75 Ω)	10 MHz to 2 GHz		0,014 + 0,081 ρ ²	method.
11.6	Signal and pulse characteristics				
11.6.3	Modulation, AM and FM: Signal generator, spectrum analyser, modulation meter, jitter meter				
i)	Amplitude Modulation (50 G	2)			
	Modulation Depth	Carrier	Modulation		
	0 % to 99 %	100 kHz to 1 MHz	20 Hz to 50 kHz	2 · 10 ⁻² · M + 0,01 %	
	0 % to 99 %	1 MHz to 1,3 GHz	20 Hz to 100 kHz	$1 \cdot 10^{-2} \cdot M + 0,01 \%$	
	1 % to 95 %	1,3 GHz to 18 GHz	60 Hz to 100 kHz	$3 \cdot 10^{-2} \cdot M$	Calibration using
	5 % to 90 %	18 GHZ to 50 GHZ	60 HZ 10 100 KHZ	3.10 ² .1M	direct measurement
ii)	Frequency Modulation (50 Ω)				detectors (e.g.
	Peak Deviation	Carrier	Modulation		spectrum analysers) /
	0 Hz to 75 kHz	100 kHz to 1 MHz	20 Hz to 35 kHz	5 ⋅ 10 ⁻³ ⋅ f + 1 Hz	calibration using
	0 Hz to 75 kHz	1 MHz to 1,3 GHz	20 Hz to 416 kHz	5 · 10 ⁻³ · f + 1 Hz	bessel functions.
	480 Hz to 3 MHz	1,3 GHz to 7 GHz	200 Hz to 1,25 MHz	5 · 10 ⁻³ · f + 10 Hz	
	480 Hz to 10 MHz	/ GHz to 18 GHz	200 Hz to 4,16 MHz	$5 \cdot 10^{-3} \cdot f + 10 \text{ Hz}$	
	480 Hz to 8 MHz	18 GHZ 10 26,5 GHZ 26 5 GHz to 50 GHz	200 HZ to 3,33 MHZ	$5 \cdot 10^{-3} \cdot t + 10 \text{ Hz}$	
	/5 KHZ [0 8 MHZ		MHz	5.10.0.1	

Original Date of Accreditation: 1980

Page 4 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor k = 2, corresponding to a confidence level of approximately 95%

ISSUED BY THE SOUTH AFRICAN NATIONAL ACCREDITATION SYSTEM

Accreditation Manager