

CERTIFICATE OF ACCREDITATION

In terms of section 22(2) (b) of the Accreditation for Conformity Assessment, Calibration and Good Laboratory Practice Act, 2006 (Act 19 of 2006), read with sections 23(1), (2) and (3) of the said Act, I hereby certify that:-

HJR MICROPTIC CC
Co. Reg. No.: 1989/032363/23

ELECTRICAL DCLF CALIBRATION LABORATORY

Accreditation Number: **CAL 022-01-00**

is a South African National Accreditation System Accredited Calibration Laboratory
provided that all SANAS conditions and requirements are complied with

This certificate is valid as per the scope as stated in the accompanying scope of accreditation
Annexure "A", bearing the above accreditation number for

DC LOW FREQUENCY METROLOGY

The facility is accredited in accordance with the recognised International Standard

ISO/IEC 17025:2017

The accreditation demonstrates technical competency for a defined scope and the operation of a
laboratory quality management system

While this certificate remains valid, the Accredited Facility named above is authorised to use the relevant
SANAS accreditation symbol to issue facility reports and/or certificates

Mr F Osman
Acting Chief Executive Officer

Effective Date: 20 June 2025
Certificate Expires: 31 August 2029

ANNEXURE A

SCOPE OF ACCREDITATION

DC LOW FREQUENCY METROLOGY

Accreditation Number: CAL 022-01-00

<u>Permanent Address of Laboratory:</u> HJR Microptic CC Electrical DCLF Calibration Laboratory 593 Smook Avenue Roseville Pretoria 0031			<u>Technical Signatories:</u> Mr AA George Mr HJ Le Roux	
<u>Postal Address:</u> PO Box 31230 Wonderboompoort 0033			<u>Nominated Representative:</u> Mr H van Deventer	
Tel: (012) 335-6663/4 Fax: (012) 335-6665 E-mail: microptic@mweb.co.za jacques@microptic.net			Issue No.: 14 Date of Issue: 20 June 2025 Expiry Date: 31 August 2029	
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT AND RANGE OF MEASURED QUANTITY	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE
1	DC Voltage (up to 1100 V, for higher voltages see 8.1)			
1.1	DC Voltage sources			
1.1.2	Low values ranges (<= 10 V)			
	0,01 mV to 1 mV 1 mV to 10 mV 10 mV to 100 mV 100 mV to 10 V	DC	8•10 ⁻³ •U + 3 µV 8•10 ⁻⁴ •U + 3 µV 2•10 ⁻⁴ •U + 3 µV 6•10 ⁻⁵ •U + 3 µV	Direct measurement or comparison with a DC voltage reference standard
1.1.3	Intermediate values (> 10 V to 1100 V)			
	10 V to 1 000 V	DC	6•10 ⁻⁵ •U + 3 µV	Direct measurement or comparison with DC Voltage source or Multi- function calibrator
1.2	DC Voltage meters			
1.2.1	Very low values (<= 1 mV)			
	0,01 mV to 1 mV	DC	8•10 ⁻³ •U + 3 µV	Direct measurement or comparison with DC voltage source and short
1.2.2	Intermediate values (> 1 mV to 1100 V)			
	1 mV to 10 mV 10 mV to 100 mV 100 mV to 1 100 V	DC	8•10 ⁻³ •U + 3 µV 8•10 ⁻⁴ •U + 3 µV 6•10 ⁻⁵ •U + 3 µV	Direct measurement or comparison with DC Voltage source or Multi- function calibrator

Original Date of Accreditation: 30 August 1998

Page 1 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%

Accreditation Manager

ANNEXURE A

Accreditation No.: CAL 022-01-00

Date of Issue: 20 June 2025

Expiry Date: 31 August 2029

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT AND RANGE OF MEASURED QUANTITY	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ($\pm$)	METHOD / PROCEDURE
2	DC resistance			
2.1	DC resistance standards and sources			
2.1.5	Multiple ranges			
	0,1 Ω to 1 Ω 1 Ω to 10 Ω 10 Ω to 10 M Ω 10 M Ω to 300 M Ω 300 M Ω to 1 G Ω	DC	$1 \cdot 10^{-2} \cdot R + 8 \text{ m}\Omega$ $6 \cdot 10^{-3} \cdot R + 8 \text{ m}\Omega$ $2 \cdot 10^{-3} \cdot R + 8 \text{ m}\Omega$ $5 \cdot 10^{-3} \cdot R$ $1 \cdot 10^{-2} \cdot R$	Direct measurement or Substitution Method.
2.2	DC resistance meters			
2.2.2	Intermediate values (> 1 Ω to 1 GΩ)			
	0,1 Ω to 1 Ω 1 Ω to 10 Ω 10 Ω to 10 M Ω 10 M Ω to 300 M Ω 300 M Ω to 1 G Ω	DC	$1 \cdot 10^{-2} \cdot R + 8 \text{ m}\Omega$ $6 \cdot 10^{-3} \cdot R + 8 \text{ m}\Omega$ $2 \cdot 10^{-3} \cdot R + 8 \text{ m}\Omega$ $5 \cdot 10^{-3} \cdot R$ $1 \cdot 10^{-2} \cdot R$	Direct measurement or Substitution Method.
3	DC current (up to 100 A, for higher currents see 8.7)			
3.1	DC current sources			
3.1.2	Intermediate values (> 0,1 mA to 20 A)			
	1 μA to 1 mA 1 mA to 1 A 1 A to 10 A 10 A to 20 A	DC	$5 \cdot 10^{-3} \cdot I$ $1 \cdot 10^{-3} \cdot I$ $3 \cdot 10^{-3} \cdot I$ $5 \cdot 10^{-3} \cdot I$	Direct measurement or Current Volt Drop Method
3.2	DC current meters			
3.2.2	Intermediate values (> 0,1 mA to 20 A)			
	1 mA to 1 A 1 A to 10 A 10 A to 20 A	DC	$1 \cdot 10^{-3} \cdot I$ $3 \cdot 10^{-3} \cdot I$ $5 \cdot 10^{-2} \cdot I$	Calibration by direct measurement
3.2.3	High values (> 20 A to 100 A)			
	20 A to 100 A (Current clamps only)	DC	$6 \cdot 10^{-2} \cdot I$	Calibration by direct measurement
4	Impedance (up to the MHz range)			
4.2	Capacitance			
4.2.4	Meters			
	330 pF to 1 mF 1 mF to 110 mF	50 Hz to 400 Hz	$5 \cdot 10^{-3} \cdot C + 20 \text{ pF}$ $3 \cdot 10^{-2} \cdot C$	Direct measurement of a reference standard capacitor or simulated through the use of a Multi-function calibrator

Original date of accreditation: 30 August 1998

Page 2 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%

Accreditation Manager

ANNEXURE A

Accreditation No.: CAL 022-01-00
Date of Issue: 20 June 2025
Expiry Date: 31 August 2029

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT AND RANGE OF MEASURED QUANTITY	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE	
5 5.2 5.2.1 5.2.2	AC voltage (up to the MHz range)				
	AC voltage up to 1000 V (for high voltage see 8.3)				
	Sources				
	3 mV to 100 mV 100 mV to 30 V	45 Hz to 10 kHz 45 Hz to 10 kHz 10 kHz to 100 kHz	5•10 ⁻³ •U 1•10 ⁻³ •U 1•10 ⁻² •U	Calibration by direct Measurement	
	30 V to 1 000 V	45 Hz to 1 kHz 1 kHz to 10 kHz	1•10 ⁻³ •U 3•10 ⁻³ •U		
	Meters				
	3 mV to 100 mV 100 mV to 30 V	45 Hz to 10 kHz 45 Hz to 10 kHz 10 kHz to 100 kHz	5•10 ⁻³ •U 1•10 ⁻³ •U 1•10 ⁻² •U	Calibration by direct Measurement	
	30 V to 1 000 V	45 Hz to 1 kHz 1 kHz to 10 kHz	1•10 ⁻³ •U 3•10 ⁻³ •U		
	6 6.2 6.2.1 6.2.2	AC current			
		AC current up to 100A (for high current see 8.6)			
Sources					
1 mA to 1 A 1 A to 10 A 10 A to 100 A		45 Hz to 1 kHz 45 Hz to 1 kHz	3•10 ⁻³ •I 5•10 ⁻³ •I 3•10 ⁻² •I	Calibration by the direct measurement or volt drop method	
Meters					
1 mA to 1 A 1 A to 10 A 10 A to 20 A 20 A to 100 A (Clamp meters only)		45 Hz to 1 kHz 45 Hz to 1 kHz 50 Hz 50 Hz	3•10 ⁻³ •I 5•10 ⁻³ •I 3•10 ⁻³ •I 3•10 ⁻² •I	Calibration by the direct measurement or volt drop method	
7 7.1 7.1.1	AC power				
	AC power and energy				
	Single phase (f <= 400 Hz)				
	1 V to 1 000 V 10 mA to 20 A	50 Hz	1•10 ⁻² •P @ PF 0,1 to 0,5 5•10 ⁻³ •P @ PF 0,5 to 0,8 3•10 ⁻³ •P @ PF 0,8 to 1	Calibration by direct comparison with a reference standard	
8 8.1 8.1.1	High voltage and current				
	High DC voltage (for voltages < 1100 V see 1)				
	High voltage sources				
	1 100 V to 30 kV	DC	1.5•10 ⁻² •U	Calibration by direct measurement or the ratio method	

Original date of accreditation: 30 August 1998

Page 3 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%

Accreditation Manager

ANNEXURE A

Accreditation No.: CAL 022-01-00
Date of Issue: 20 June 2025
Expiry Date: 31 August 2029

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ($\pm$)	METHOD / PROCEDURE
8.3	High AC voltage and voltage transformers (for voltages <1100V see 5.2 and 5.3)			
8.3.2	Sources			
	1 kV to 25 kV	50 Hz	$4 \cdot 10^{-2} \cdot U$	Calibration by direct measurement or the Ratio Method.
8.6	High AC current and current transformers (for currents ≤ 100 A see 6.2 and 6.3)			
8.6.2	Meters			
	100 A to 1 000 A (Clamp meters only)	50 Hz	$5 \cdot 10^{-3} \cdot I$	Calibration by direct measurement or the Ratio Method.
8.7	High DC current (for currents ≤ 100 A see 3, for shunts see 2.1.4)			
8.7.2	Meters			
	100 A to 1 000 A (Clamp Meters only)	DC	$5 \cdot 10^{-3} \cdot I$	Calibration by direct measurement or the Ratio Method.
A	Thermometry equipment (Electrical Simulation)			
A1	Sources			
	Voltage	- 200 °C to + 700 °C + 700 °C to + 2 000 °C	0,5 K 1,0 K	Calibration by the measurement of the voltage or resistance equivalent to the thermometer type.
	Resistance	- 200 °C to + 600 °C	0,9 K	
A2	Meters			
	Voltage	-200 °C to +700 °C +700 °C to + 2 000 °C	0,5 K 1,0 K	Calibration by the sourcing of voltages or resistance equivalent to the thermometer type.
	Resistance	- 200 °C to + 600 °C	0,9 K	
C	Frequency			
	Sinusoidal	1 Hz to 10 MHz	$3 \cdot 10^{-4} \cdot f + 10$ mHz	Calibration by the direct measurement of frequency source.
Z	On-site calibrations for items 1,2,3,4,5,6,7,8, and A and C above			

Original date of accreditation: 30 August 1998

Page 4 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%

ISSUED BY THE SOUTH AFRICAN NATIONAL ACCREDITATION SYSTEM

Accreditation Manager