

# **CERTIFICATE OF ACCREDITATION**

*In terms of section 22(2) (b) of the Accreditation for Conformity Assessment, Calibration and Good Laboratory Practice Act, 2006 (Act 19 of 2006), read with sections 23(1), (2) and (3) of the said Act, I hereby certify that:-*

**TESTO SOUTH AFRICA (PTY) LTD**  
**Co. Reg. No.: 2015/403399/07**

**GAS CALIBRATION LABORATORY**  
**KEMPTON PARK**

**Accreditation Number: CAL 076-15-00**

is a South African National Accreditation System Accredited Calibration laboratory  
provided that all SANAS conditions and requirements are complied with

This certificate is valid as per the scope as stated in the accompanying scope of accreditation  
Annexure "A", bearing the above accreditation number for

## **GAS METROLOGY**

The facility is accredited in accordance with the recognised International Standard

**ISO/IEC 17025:2017**

The accreditation demonstrates technical competency for a defined scope and the operation of a  
laboratory quality management system

While this certificate remains valid, the Accredited Facility named above is authorised to use the  
relevant SANAS accreditation symbol to issue facility reports and/or certificates

---

**Mr F Osman**  
**Acting Chief Executive Officer**

**Effective Date: 24 August 2025**  
**Certificate Expires: 23 August 2030**

## ANNEXURE A

### SCOPE OF ACCREDITATION

#### GAS METROLOGY

Accreditation Number: CAL 076-15-00

<b>Permanent Address of Laboratory:</b> Testo South Africa (Pty) Ltd Gas Calibration Laboratory Unit 1 Gleneagle Office Park Cnr Braambos and Monument Road Glen Marais Kempton Park 1619		<b>Technical Signatory:</b> Mr M Talane	
<b>Postal Address:</b> Suite 42, Private Bag 7 Aston Manor Kempton Park		<b>Nominated Representative:</b> Mr J Kirkland	
Tel: (011) 380-8060 Cell: 066 476 4229 E-mail: <a href="mailto:jkirkland@testo.co.za">jkirkland@testo.co.za</a>	Issue No.: 06 Date of Issue: 24 August 2025 Expiry Date: 23 August 2030		
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )
4	<b>GASES</b>		
4.4	<b>Calibration of Gas Analysers</b>		
4.4.1	Oxygen (O <sub>2</sub> in N <sub>2</sub> )	0 to 3 % mol/mol 3 to 18 % mol/mol	0,03 % mol/mol 0,10 % mol/mol
4.4.2	Carbon Monoxide (CO in N <sub>2</sub> )	0 to 100 $\mu$ mol/mol 100 to 1 000 $\mu$ mol/mol	2,0 % 1,0 %
4.4.3	Nitrogen Monoxide (NO in N <sub>2</sub> )	0 to 100 $\mu$ mol/mol 100 to 1 000 $\mu$ mol/mol	2,0 % 1,0 %
4.4.4	Nitrogen Dioxide (NO <sub>2</sub> in N <sub>2</sub> )	0 to 10 $\mu$ mol/mol 10 to 100 $\mu$ mol/mol	2,0 % 1,0 %
4.4.5	Sulphur Dioxide (SO <sub>2</sub> in N <sub>2</sub> )	0 to 100 $\mu$ mol/mol 100 to 1 000 $\mu$ mol/mol	2,0 % 1,0 %
4.4.6	Carbon Dioxide (CO <sub>2</sub> in N <sub>2</sub> )	0 to 5 % mol/mol 5 to 15 % mol/mol 15 to 40 % mol/mol	0,1 % mol/mol 0,2 % mol/mol 0,6 % mol/mol

Original Date of Accreditation: 21 February 2019

Page 1 of 1

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor  $k = 2$ , corresponding to a confidence level of approximately 95%

ISSUED BY THE SOUTH AFRICAN NATIONAL ACCREDITATION SYSTEM

**Accreditation Manager**